

CERTIFICATE OF COMPLIANCE

Certificate Number 20150223-E90509
Report Reference E90509-20080926
Issue Date 2015-FEBRUARY-23

Issued to: Schneider Electric Industries SAS
35 Rue Joseph Monier
92506 Rueil Malmaison
FRANCE

**This is to certify that
representative samples of**

COMPONENT - PROTECTORS, SUPPLEMENTARY
Component - Supplementary Protector,
Type C60H-DC ; followed by ratings 0.5, 1, 2, 3, 4, 5, 6, 8,
10, 13 15, 16, 20, 25, 30, 32, 35, 40, 50, 63 followed by B,
C or K

Have been investigated by UL in accordance with the
Standard(s) indicated on this Certificate.

Standard(s) for Safety: UL1077 - Supplementary Protectors for Use in Electrical
Equipment

Additional Information: See the UL Online Certifications Directory at
www.ul.com/database for additional information

Only those products bearing the UL Certification Mark should be considered as being covered by UL's
Certification and Follow-Up Service.

Recognized components are incomplete in certain constructional features or restricted in performance
capabilities and are intended for use as components of complete equipment submitted for investigation rather
than for direct separate installation in the field. The final acceptance of the component is dependent upon its
installation and use in complete equipment submitted to UL LLC.

Look for the UL Certification Mark on the product.



Bruce Mahrenholz, Assistant Chief Engineer, Global Inspection and Field Services

UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please
contact a local UL Customer Service Representative at www.ul.com/contactus

